



PERRY JOHNSON LABORATORY ACCREDITATION, INC.

Certificate of Accreditation

Perry Johnson Laboratory Accreditation, Inc. has assessed the Laboratory of:

PCI

8100 Brownleigh Dr., #100A, Raleigh, NC 27617

(Hereinafter called the Organization) and hereby declares that Organization is accredited in accordance with the recognized International Standard:

ISO/IEC 17025:2017

This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (as outlined by the joint ISO-ILAC-IAF Communiqué dated April 2017):

Dimensional, Electrical, Mechanical, Chemical, Time & Frequency, and Thermodynamic Calibration
(As detailed in the supplement)

Accreditation claims for such testing and/or calibration services shall only be made from addresses referenced within this certificate. This Accreditation is granted subject to the system rules governing the Accreditation referred to above, and the Organization hereby covenants with the Accreditation body's duty to observe and comply with the said rules.

For PJLA:

Initial Accreditation Date:

September 12, 2011

Issue Date:

November 26, 2019

Expiration Date:

January 31, 2022

Accreditation No.:

54363

Certificate No.:

L19-586

Tracy Szerszen
President/Operations Manager

Perry Johnson Laboratory
Accreditation, Inc. (PJLA)
755 W. Big Beaver, Suite 1325
Troy, Michigan 48084

The validity of this certificate is maintained through ongoing assessments based on a continuous accreditation cycle. The validity of this certificate should be confirmed through the PJLA website: www.pjilabs.com



Certificate of Accreditation: Supplement

PCI

8100 Brownleigh Dr., #100A, Raleigh, NC 27617
 Contact Name: Frank Reagan Phone: 919-781-7787

Accreditation is granted to the facility to perform the following calibrations:

Dimensional

| MEASURED INSTRUMENT, QUANTITY OR GAUGE | RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE | CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm) | CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED |
|---|---|--|--|
| Equipment to Measure Length ^{FO} | 0.001 in to 0.4 in | 21 μ in/in + 30 μ in | Gage Block Set |
| | 0.401 in to 1 in | 48 μ in/in + 19 μ in | |
| | 1.001 in to 7 in | 59 μ in/in + 8.4 μ in | |
| | 7.001 in to 24 in | 57 μ in/in + 140 μ in | Distance Bars |

Electrical

| MEASURED INSTRUMENT, QUANTITY OR GAUGE | RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE | CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm) | CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED |
|---|---|--|--|
| Equipment to Measure DC Voltage ^{FO} | 0.01 mV to 329.999 9 mV | 0.001 3 mV + 0.000 023 mV/mV | Fluke 5520A |
| | 0.33 V to 3.299 999 V | 0.000 58 V + 0.000 005 4 V/V | |
| | 3.3 V to 32.999 99 V | 0.000 58 V + 0.000 005 5 V/V | |
| | 33 V to 329.999 9 V | 0.000 32 V + 0.000 021 V/V | |
| | 330 V to 1 020 V | 0.001 7 V + 0.000 021 V/V | |
| Equipment to Output DC Voltage ^{FO} | 0.000 3 μ V to 200 mV | 5 μ V/V + 0.1 μ V | Fluke 8508A Reference Multi-meter |
| | 200 mV to 2 V | 3.5 μ V/V + 0.4 μ V | |
| | 2 V to 20 V | 3.5 μ V/V + 4 μ V | |
| | 20 V to 200 V | 5.5 μ V/V + 0.04 mV | Fluke 80K-6 HV Probe with 8060A Multi-meter |
| | 200 V to 1 000 V | 5.5 μ V/V + 0.5 mV | |
| | 1 000 V to 6 000 V | 10 mV/V + 2 V | |
| Equipment to Measure DC Current ^{FO} | 0.001 μ A to 329.999 μ A | 0.023 μ A + 0.000 18 μ A/A | Fluke 5520A |
| | 0.33 mA to 3.299 99 mA | 0.000 084 mA + 0.000 11 mA/mA | |
| | 3.3 mA to 32.999 9 mA | 0.000 65 mA + 0.000 11 mA/mA | |
| | 33 mA to 329.999 mA | 0.003 mA + 0.000 11 mA/mA | |
| | 0.33 mA to 1.099 99 mA | 0.000 047 A + 0.000 24 A/A | |
| | 1.1 ADC to 2.999 99 ADC | 0.000 041 + 0.000 45 A/A | |
| | 3 A to 10.999 9 A | 0.000 83 A + 0.000 56 A/A | |
| | 11 A to 20.5 A | 0.002 6 A + 0.001 2 A/A | |
| Equipment to Output DC Current ^{FO} | 0.001 2 μ A to 200 μ A | 12 μ A/A + 0.4 nA | Fluke 8508A Reference Multi-meter |
| | 200 μ A to 2 mA | 12 μ A/A + 4 nA | |
| | 2 mA to 20 mA | 14 μ A/A + 0.04 μ A | |
| | 20 mA to 200 mA | 48 μ A/A + 0.8 μ A | |
| | 200 mA to 2 A | 185 μ A/A + 16 μ A | |
| | 2 A to 20 A | 400 μ A/A + 400 μ A | |



Certificate of Accreditation: Supplement

PCI

8100 Brownleigh Dr., #100A, Raleigh, NC 27617
 Contact Name: Frank Reagan Phone: 919-781-7787

Accreditation is granted to the facility to perform the following calibrations:

Electrical

| MEASURED INSTRUMENT, QUANTITY OR GAUGE | RANGE (AND SPECIFICATION WHERE APPROPRIATE) | CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm) | CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED |
|---|---|--|---|
| Equipment to Measure Resistance- ^{FO} | 0.001 Ω + 10.999 9 Ω | 0.001 2 Ω + 0.000 047 Ω/Ω | Fluke 5520A |
| | 11 Ω to 32.999 9 Ω | 0.001 7 Ω + 0.000 035 Ω/Ω | |
| | 33 Ω to 109.999 9 Ω | 0.006 1 Ω + 0.000 27 Ω/Ω | |
| | 110 Ω to 329.999 9 Ω | 0.002 3 Ω + 0.000 033 Ω/Ω | |
| | 0.33 k Ω to 1.099 999 k Ω | 0.000 003 9 k Ω + 0.000 032 k Ω/Ω | |
| | 1.1 k Ω to 3.299 999 k Ω | 0.000 053 k Ω + 0.000 017 k $\Omega/k\Omega$ | |
| | 3.3 k Ω to 10.999 99 k Ω | 0.000 55 k Ω + 0.000 014 k $\Omega/k\Omega$ | |
| | 11 k Ω to 32.999 99 k Ω | 0.000 52 k Ω + 0.000 029 k $\Omega/k\Omega$ | |
| | 33 k Ω to 109.999 9 k Ω | 0.000 39 k Ω + 0.000 033 k $\Omega/k\Omega$ | |
| | 110 k Ω to 329.999 9 k Ω | 0.005 k Ω + 0.000 033 k $\Omega/k\Omega$ | |
| | 330 k Ω to 1.099 999 M Ω | 0.000 003 8 M Ω + 0.000 037 M $\Omega/M\Omega$ | |
| | 1.1 M Ω to 3.299 999 M Ω | 0.000 051 M Ω + 0.000 068 M $\Omega/M\Omega$ | |
| | 3.3 M Ω to 10.999 99 M Ω | 0.000 37 M Ω + 0.000 14 M $\Omega/M\Omega$ | |
| | 11 M Ω to 32.999 99 M Ω | 0.002 8 M Ω + 0.000 3 M $\Omega/M\Omega$ | |
| | 33 M Ω to 109.999 9 M Ω | 0.003 5 M Ω + 0.000 59 M $\Omega/M\Omega$ | |
| 110 M Ω to 329.999 9 M Ω | 0.12 M Ω + 0.003 5 M $\Omega/M\Omega$ | | |
| 330 M Ω to 1 100 M Ω | 0.58 M Ω + 0.018 M $\Omega/M\Omega$ | | |
| Equipment to Output Resistance ^{FO} | 0.000 012 Ω to 2 Ω | 17 $\mu\Omega/\Omega$ + 4 $\mu\Omega$ | Fluke 8508A Reference Multi- meter |
| | 2 Ω to 20 Ω | 9.5 $\mu\Omega/\Omega$ + 28 $\mu\Omega$ | |
| | 20 Ω to 200 Ω | 8 $\mu\Omega/\Omega$ + 50 $\mu\Omega$ | |
| | 200 Ω to 2 k Ω | 8 $\mu\Omega/\Omega$ + 500 $\mu\Omega$ | |
| | 2 k Ω to 20 k Ω | 8 $\mu\Omega/\Omega$ + 5.0 m Ω | |
| | 20 k Ω to 200 k Ω | 8 $\mu\Omega/\Omega$ + 50 m Ω | |
| | 200 k Ω to 2 M Ω | 9 $\mu\Omega/\Omega$ + 1 Ω | |
| | 2 M Ω to 20 M Ω | 20 $\mu\Omega/\Omega$ + 100 Ω | |
| | 20 M Ω to 200 M Ω | 120 $\mu\Omega/\Omega$ + 10 k Ω | |
| | 200 M Ω to 2 G Ω | 1 510 $\mu\Omega/\Omega$ + 100 k Ω | |
| | 2 G Ω to 20 G Ω | 1 510 $\mu\Omega/\Omega$ + 1 M Ω | |



Certificate of Accreditation: Supplement

PCI

8100 Brownleigh Dr., #100A, Raleigh, NC 27617
 Contact Name: Frank Reagan Phone: 919-781-7787

Accreditation is granted to the facility to perform the following calibrations:

Electrical

| MEASURED INSTRUMENT, QUANTITY OR GAUGE | RANGE (AND SPECIFICATION WHERE APPROPRIATE) | CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm) | CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED |
|--|---|--|---|
| Equipment to Measure Capacitance ^{FO} | 0.19 nF to 0.399 9 nF | 0.11 nF + 0.009 1 nF/nF | Fluke 5520A |
| | 0.4 nF to 1.099 9 nF | 0.011 nF+ 0.006 6 nF/nF | |
| | 1.1 nF to 3.299 9 nF | 0.012 nF + 0.005 9 nF/nF | |
| | 3.3 nF to 10.999 9 nF | 0.012 nF + 0.002 9 nF/nF | |
| | 11 nF to 32.999 9 nF | 0.12 nF + 0.003 nF/nF | |
| | 33 nF to 109.999 nF | 0.12 nF + 0.002 9 nF/nF | |
| | 110 nF to 329.999 nF | 0.33 nF + 0.003 1 nF/nF | |
| | 0.33 μ F to 1.099 99 μ F | 0.001 3 μ F + 0.002 9 μ F/ μ F | |
| | 1.1 μ F to 3.299 99 μ F | 0.003 4 μ F + 0.003 μ F/ μ F | |
| | 3.3 μ F to 10.999 9 μ F | 0.011 μ F + 0.003 1 μ F/ μ F | |
| | 11 μ F to 32.999 9 μ F | 0.33 μ F + 0.004 8 μ F/ μ F | |
| | 33 μ F to 109.99 μ F | 0.11 μ F + 0.005 5 μ F/ μ F | |
| | 110 μ F to 329.99 μ F | 0.34 μ F + 0.005 4 μ F/ μ F | |
| | 0.33 mF to 1.099 99 mF | 0.001 3 mF + 0.005 2 mF/mF | |
| | 1.1 mF to 3.299 9 mF | 0.003 5 mF + 0.005 2 mF/mF | |
| | 3.3 mF to 10.999 9 mF | 0.12 mF + 0.005 3 mF/mF | |
| 11 mF to 32.999 9 mF | 0.034 mF + 0.008 7 mF/mF | | |
| 33 mF to 110 mF | 0.12 mF + 0.013 mF/mF | | |
| Equipment to Measure AC Voltage (at the listed frequencies) ^{FO} | | | |
| 10 Hz to 45 Hz | 1 mV to 32.999 mV | 0.008 9 mV + 0.000 878 mV/mV | |
| 45 Hz to 10 kHz | 1 mV to 32.999 mV | 0.009 mV + 0.000 148 mV/mV | |
| 10 kHz to 20 kHz | 1 mV to 32.999 mV | 0.009 mV + 0.000 202 mV/mV | |
| 20 kHz to 50 kHz | 1 mV to 32.999 mV | 0.008 8 mV + 0.001 11 mV/mV | |
| 50 kHz to 100 kHz | 1 mV to 32.999 mV | 0.015 mV + 0.004 02 mV/mV | |
| 100 kHz to 500 kHz | 1 mV to 32.999 mV | 0.058 mV + 0.009 23 mV | |
| Equipment to Measure AC Voltage (at the listed frequencies) ^{FO} | | | |
| 10 Hz to 45 Hz | 33 mV to 329.999 mV | 0.009 2 mV + 0.003 46 mV/mV | |
| 45 Hz to 10 kHz | 33 mV to 329.999 mV | 0.009 mV + 0.001 74 mV/mV | |
| 10 kHz to 20 kHz | 33 mV to 329.999 mV | 0.009 mV + 0.002 23 mV/mV | |
| 20 kHz to 50 kHz | 33 mV to 329.999 mV | 0.009 2 mV + 0.004 04 mV/mV | |
| 50 kHz to 100 kHz | 33 mV to 329.999 mV | 0.037 mV + 0.000 924 mV/mV | |
| 100 kHz to 500 kHz | 33 mV to 329.999 mV | 0.081 mV + 0.002 31 mV/mV | |



Certificate of Accreditation: Supplement

PCI

8100 Brownleigh Dr., #100A, Raleigh, NC 27617
 Contact Name: Frank Reagan Phone: 919-781-7787

Accreditation is granted to the facility to perform the following calibrations:

Electrical

| MEASURED INSTRUMENT, QUANTITY OR GAUGE | RANGE (AND SPECIFICATION WHERE APPROPRIATE) | CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm) | CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED |
|--|---|--|---|
| Equipment to Measure AC Voltage (at the listed frequencies) ^{FO} | | | Fluke 5520A |
| 10 Hz to 45 Hz | 0.33 V to 3.299 99 V | 0.000 058 V + 0.000 346 V/V | |
| 45 Hz to 10 kHz | 0.33 V to 3.299 99 V | 0.000 069 V + 0.000 173 V/V | |
| 10 kHz to 20 kHz | 0.33 V to 3.299 99 V | 0.000 069 V + 0.000 219 V/V | |
| 20 kHz to 50 kHz | 0.33 V to 3.299 99 V | 0.000 076 V + 0.000 292 V/V | |
| 50 kHz to 100 kHz | 0.33 V to 3.299 99 V | 0.00 14 V + 0.000 808 V/V | |
| 100 kHz to 500 kHz | 0.33 V to 3.299 99 V | 0.000 69 V + 0.002 77 V/V | |
| Equipment to Measure AC Voltage (at the listed frequencies) ^{FO} | | | |
| 10 Hz to 45 Hz | 3.3 V to 32.999 9 V | 0.000 75 V + 0.003 46 V/V | |
| 45 Hz to 10 kHz | 3.3 V to 32.999 9 V | 0.000 69 V + 0.000 173 V/V | |
| 10 kHz to 20 kHz | 3.3 V to 32.999 9 V | 0.000 69 V + 0.000 277 V/V | |
| 20 kHz to 50 kHz | 3.3 V to 32.999 9 V | 0.000 69 V + 0.000 404 V/V | |
| 50 kHz to 100 kHz | 3.3 V to 32.999 9 V | 0.001 8 V + 0.001 04 V/V | |
| Equipment to Measure AC Voltage (at the listed frequencies) ^{FO} | | | |
| 45 Hz to 1 kHz | 33 V to 329.999 V | 0.002 3 V + 0.000 219 V/V | |
| 1 kHz to 10 kHz | 33 V to 329.999 V | 0.006 9 V + 0.000 231 V/V | |
| 10 kHz to 20 kHz | 33 V to 329.999 V | 0.006 9 V + 0.000 289 V/V | |
| 20 kHz to 50 kHz | 33 V to 329.999 V | 0.006 9 V + 0.000 346 V/V | |
| 50 kHz to 100 kHz | 33 V to 329.999 V | 0.058 V + 0.002 31 V/V | |
| Equipment to Measure AC Voltage (at the listed frequencies) ^{FO} | | | |
| 45 Hz to 1 kHz | 330 V to 1 020 V | 0.12 V + 0.000 346 V/V | |
| 1 kHz to 5 kHz | 330 V to 1 020 V | 0.04 V + 0.000 261 V/V | |
| 5 kHz to 10 kHz | 330 V to 1 020 V | 0.012 V + 0.000 346 V/V | |
| Equipment to Output AC Voltage (at the listed frequencies) ^{FO} | | | Fluke 8508A Reference Multi-meter |
| 1 Hz to 10 Hz | 0.052 mV to 200 mV | 165 μ V/V + 14 μ V | |
| 10 Hz to 40 Hz | 0.052 mV to 200 mV | 140 μ V/V + 4 μ V | |
| 40 Hz to 100 Hz | 0.052 mV to 200 mV | 115 μ V/V + 4 μ V | |
| 100 Hz to 2 kHz | 0.052 mV to 200 mV | 110 μ V/V + 2 μ V | |
| 2 kHz to 10 kHz | 0.052 mV to 200 mV | 135 μ V/V + 4 μ V | |
| 10 kHz to 30 kHz | 0.052 mV to 200 mV | 340 μ V/V + 8 μ V | |
| 30 kHz to 100 kHz | 0.052 mV to 200 mV | 765 μ V/V + 20 μ V | |



Certificate of Accreditation: Supplement

PCI

8100 Brownleigh Dr., #100A, Raleigh, NC 27617
 Contact Name: Frank Reagan Phone: 919-781-7787

Accreditation is granted to the facility to perform the following calibrations:

Electrical

| MEASURED INSTRUMENT, QUANTITY OR GAUGE | RANGE (AND SPECIFICATION WHERE APPROPRIATE) | CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm) | CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED |
|---|---|--|---|
| Equipment to Output AC Voltage (at the listed frequencies) ^{FO} | | | Fluke 8508A Reference Multi-meter |
| 1 Hz to 10 Hz | 200 mV to 2 V | 150 μ V/V + 120 μ V | |
| 10 Hz to 40 Hz | 200 mV to 2 V | 115 μ V/V + 20 μ V | |
| 40 Hz to 100 Hz | 200 mV to 2 V | 90 μ V/V + 20 μ V | |
| 100 Hz to 2 kHz | 200 mV to 2 V | 75 μ V/V + 20 μ V | |
| 2 kHz to 10 kHz | 200 mV to 2 V | 110 μ V/V + 20 μ V | |
| 10 kHz to 30 kHz | 200 mV to 2 V | 220 μ V/V + 40 μ V | |
| 30 kHz to 100 kHz | 200 mV to 2 V | 570 μ V/V + 200 μ V | |
| 100 kHz to 300 kHz | 200 mV to 2 V | 3 mV/V + 2 mV | |
| 300 kHz to 1 MHz | 200 mV to 2 V | 10 mV/V + 20 mV | |
| Equipment to Output AC Voltage (at the listed frequencies) ^{FO} | | | Fluke 8508A Reference Multi-meter |
| 1 Hz to 10 Hz | 2 V to 20 V | 150 μ V/V + 1.2 mV | |
| 10 Hz to 40 Hz | 2 V to 20 V | 115 μ V/V + 200 μ V | |
| 40 Hz to 100 Hz | 2 V to 20 V | 90 μ V/V + 200 μ V | |
| 100 Hz to 2 kHz | 2 V to 20 V | 75 μ V/V + 200 μ V | |
| 2 kHz to 10 kHz | 2 V to 20 V | 110 μ V/V + 200 μ V | |
| 10 kHz to 30 kHz | 2 V to 20 V | 220 μ V/V + 400 μ V | |
| 30 kHz to 100 kHz | 2 V to 20 V | 570 μ V/V + 2 mV | |
| 100 kHz to 300 kHz | 2 V to 20 V | 3 mV/V + 20 mV | |
| 300 kHz to 1 MHz | 2 V to 20 V | 10 mV/V + 200 mV | |
| Equipment to Output AC Voltage (at the listed frequencies) ^{FO} | | | Fluke 8508A Reference Multi-meter |
| 1 Hz to 10 Hz | 20 V to 200 V | 150 μ V/V + 12 mV | |
| 10 Hz to 40 Hz | 20 V to 200 V | 115 μ V/V + 2 mV | |
| 40 Hz to 100 Hz | 20 V to 200 V | 90 μ V/V + 2 mV | |
| 100 Hz to 2 kHz | 20 V to 200 V | 75 μ V/V + 2 mV | |
| 2 kHz to 10 kHz | 20 V to 200 V | 110 μ V/V + 2 mV | |
| 10 kHz to 30 kHz | 20 V to 200 V | 220 μ V/V + 4 mV | |
| 30 kHz to 100 kHz | 20 V to 200 V | 570 μ V/V + 20 mV | |
| 100 kHz to 300 kHz | 20 V to 200 V | 3 mV/V + 200 mV | |
| 300 kHz to 1 MHz | 20 V to 200 V | 10 mV/V + 2 mV | |



Certificate of Accreditation: Supplement

PCI

8100 Brownleigh Dr., #100A, Raleigh, NC 27617
 Contact Name: Frank Reagan Phone: 919-781-7787

Accreditation is granted to the facility to perform the following calibrations:

Electrical

| MEASURED INSTRUMENT, QUANTITY OR GAUGE | RANGE (AND SPECIFICATION WHERE APPROPRIATE) | CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm) | CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED |
|--|---|--|---|
| Equipment to Output AC Voltage (at the listed frequencies) ^{FO} | | | Fluke 8508A Reference Multi-meter |
| 1 Hz to 10 Hz | 200 to 1 050 V | 150 μ V/V + 73.5 mV | |
| 10 Hz to 40 Hz | 200 V to 1 050 V | 120 μ V/V + 21 mV | |
| 40 Hz to 10 kHz | 200 V to 1 050 V | 115 μ V/V + 21 mV | |
| 10 kHz to 30 kHz | 200 V to 1 050 V | 225 μ V/V + 42 mV | |
| 30 kHz to 100 kHz | 200 V to 1 050 V | 580 μ V/V + 210 mV | |
| 45 Hz to 1 kHz | 1 000 V to 6 000 V | 11.2 mV/V + 12 V | Fluke 80K-6 HV Probe with 8060A Multi-meter |
| Equipment to Measure AC Current (at the listed frequencies) ^{FO} | | | Fluke 5520A |
| 10 Hz to 20 Hz | 29 μ A to 329.99 μ A | 0.12 μ A + 0.002 31 μ A/ μ A | |
| 20 Hz to 45 Hz | 29 μ A to 329.99 μ A | 0.12 μ A + 0.001 73 μ A/ μ A | |
| 45 Hz to 1 kHz | 29 μ A to 329.99 μ A | 0.12 μ A + 0.001 44 μ A/ μ A | |
| 1 kHz to 5 kHz | 29 μ A to 329.99 μ A | 0.17 μ A + 0.003 46 μ A/ μ A | |
| 5 kHz to 10 kHz | 29 μ A to 329.99 μ A | 0.23 μ A + 0.009 24 μ A/ μ A | |
| 10 kHz to 30 kHz | 29 μ A to 329.99 μ A | 0.46 μ A + 0.018 5 μ A/ μ A | |
| Equipment to Measure AC Current (at the listed frequencies) ^{FO} | | | |
| 10 Hz to 20 Hz | 0.33 mA to 3.299 9 mA | 0.000 17 mA + 0.002 31 mA/mA | |
| 20 Hz to 45 Hz | 0.33 mA to 3.299 9 mA | 0.000 17 mA + 0.001 44 mA/mA | |
| 45 Hz to 1 kHz | 0.33 mA to 3.299 9 mA | 0.000 17 mA + 0.001 15 mA/mA | |
| 1 kHz to 5 kHz | 0.33 mA to 3.299 9 mA | 0.000 23 mA + 0.002 31 mA/mA | |
| 5 kHz to 10 kHz | 0.33 mA to 3.299 9 mA | 0.000 35 mA + 0.005 77 mA/mA | |
| 10 kHz to 30 kHz | 0.33 mA to 3.299 9 mA | 0.000 69 mA + 0.011 5 mA/mA | |
| Equipment to Measure AC Current (at the listed frequencies) ^{FO} | | | |
| 10 Hz to 20 Hz | 3.3 mA to 32.999 mA | 0.002 3 mA + 0.002 08 mA/mA | |
| 20 Hz to 45 Hz | 3.3 mA to 32.999 mA | 0.002 3 mA + 0.001 04 mA/mA | |
| 45 Hz to 1 kHz | 3.3 mA to 32.999 mA | 0.002 3 mA + 0.000 462 mA/mA | |
| 1 kHz to 5 kHz | 3.3 mA to 32.999 mA | 0.002 3 mA + 0.000 924 mA/mA | |
| 5 kHz to 10 kHz | 3.3 mA to 32.999 mA | 0.003 5 mA + 0.002 31 mA/mA | |
| 10 kHz to 30 kHz | 3.3 mA to 32.999 mA | 0.004 6 mA + 0.004 62 mA/mA | |



Certificate of Accreditation: Supplement

PCI

8100 Brownleigh Dr., #100A, Raleigh, NC 27617
 Contact Name: Frank Reagan Phone: 919-781-7787

Accreditation is granted to the facility to perform the following calibrations:

Electrical

| MEASURED INSTRUMENT, QUANTITY OR GAUGE | RANGE (AND SPECIFICATION WHERE APPROPRIATE) | CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm) | CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED |
|--|---|--|---|
| Equipment to Measure AC Current (at the listed frequencies) ^{F0} | | | Fluke 5520A |
| 10 Hz to 20 Hz | 33 mA to 329.99 mA | 0.023 mA + 0.002 08 mA/mA | |
| 20 Hz to 45 Hz | 33 mA to 329.99 mA | 0.023 mA + 0.001 04 mA/mA | |
| 45 Hz to 1 kHz | 33 mA to 329.99 mA | 0.023 mA + 0.000 462 mA/mA | |
| 1 kHz to 5 kHz | 33 mA to 329.99 mA | 0.023 mA + 0.000 923 mA/mA | |
| 5 kHz to 10 kHz | 33 mA to 329.99 mA | 0.12 mA + 0.002 31 mA/mA | |
| 10 kHz to 30 kHz | 33 mA to 329.99 mA | 0.23 mA + 0.004 62 mA/mA | |
| Equipment to Measure AC Current (at the listed frequencies) ^{F0} | | | |
| 10 Hz to 45 Hz | 0.33 A to 1.099 99 A | 0.000 12 A + 0.002 08 A/A | |
| 45 Hz to 1 kHz | 0.33 A to 1.099 99 A | 0.000 12 A + 0.000 573 A/A | |
| 1 kHz to 5 kHz | 0.33 A to 1.099 99 A | 0.001 2 A + 0.006 93 A/A | |
| 5 kHz to 10 kHz | 0.33 A to 1.099 99 A | 0.005 8 A + 0.028 9 A/A | |
| Equipment to Measure AC Current (at the listed frequencies) ^{F0} | | | |
| 10 Hz to 45 Hz | 1.1 A to 2.999 99 A | 0.000 21 A + 0.002 06 A/A | |
| 45 Hz to 1 kHz | 1.1 A to 2.999 99 A | 0.000 35 A + 0.000 641 A/A | |
| 1 kHz to 5 kHz | 1.1 A to 2.999 99 A | 0.001 2 A + 0.006 93 A/A | |
| 5 kHz to 10 kHz | 1.1 A to 2.999 99 A | 0.005 8 A + 0.028 9 A/A | |
| Equipment to Measure AC Current (at the listed frequencies) ^{F0} | | | |
| 10 Hz to 45 Hz | 3 A to 10.999 9 A | 0.002 4 A + 0.000 69 A/A | |
| 45 Hz to 1 kHz | 3 A to 10.999 9 A | 0.002 3 A + 0.001 15 A/A | |
| 1 kHz to 5 kHz | 3 A to 10.999 9 A | 0.002 3 A + 0.034 6 A/A | |
| Equipment to Measure AC Current (at the listed frequencies) ^{F0} | | | |
| 45 Hz to 100 Hz | 11 A to 20.5 A | 0.005 8 A + 0.001 39 A/A | |
| 100 Hz to 1 kHz | 11 A to 20.5 A | 0.005 8 A + 0.001 73 A/A | |
| 1 kHz to 5 kHz | 11 A to 20.5 A | 0.005 8 A + 0.034 6 A/A | |
| Equipment to Output AC Current (at the listed frequencies) ^{F0} | | | Fluke 8508A Reference Multi-meter |
| 1 Hz to 10 Hz | 0.06 μ A to 199.99 μ A | 500 μ A/A + 0.02 μ A | |
| 10 Hz to 10 kHz | 0.06 μ A to 199.99 μ A | 500 μ A/A + 0.02 μ A | |
| 10 kHz to 30 kHz | 0.06 μ A to 199.99 μ A | 710 μ A/A + 0.02 μ A | |
| 30 kHz to 100 kHz | 0.06 μ A to 199.99 μ A | 4 mA/A + 0.02 μ A | |



Certificate of Accreditation: Supplement

PCI

8100 Brownleigh Dr., #100A, Raleigh, NC 27617
 Contact Name: Frank Reagan Phone: 919-781-7787

Accreditation is granted to the facility to perform the following calibrations:

Electrical

| MEASURED INSTRUMENT, QUANTITY OR GAUGE | RANGE (AND SPECIFICATION WHERE APPROPRIATE) | CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm) | CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED |
|---|---|--|--|
| Equipment to Output AC Current (at the listed frequencies) ^{FO} | | | Fluke 8508A Reference Multi-meter |
| 1 Hz to 10 Hz | 200 μ A to 1.999 900 mA | 310 μ A/A + 0.2 μ A | |
| 10Hz to 10 kHz | 200 μ A to 1.999 900 mA | 300 μ A/A + 0.2 μ A | |
| 10 kHz to 30 kHz | 200 μ A to 1.999 900 mA | 710 μ A/A + 0.2 μ A | |
| 30 kHz to 100 kHz | 200 μ A to 1.999 900 mA | 4 mA/A + 0.2 μ A | |
| Equipment to Output AC Current (at the listed frequencies) ^{FO} | | | |
| 1 Hz to 10 Hz | 2 mA to 19.999 900 mA | 310 μ A/A + 2 μ A | |
| 10Hz to 10 kHz | 2 mA to 19.999 900 mA | 300 μ A/A + 2 μ A | |
| 10 kHz to 30 kHz | 2 mA to 19.999 900 mA | 710 μ A/A + 2 μ A | |
| 30 kHz to 100 kHz | 2 mA to 19.999 900 mA | 4 mA/A + 2 μ A | |
| Equipment to Output AC Current (at the listed frequencies) ^{FO} | | | |
| 1 Hz to 10 Hz | 20 mA to 199.990 0 mA | 310 μ A/A + 20 μ A | |
| 10Hz to 10 kHz | 20 mA to 199.990 0 mA | 290 μ A/A + 20 μ A | |
| 10 kHz to 30 kHz | 20 mA to 199.990 0 mA | 625 μ A/A + 20 μ A | |
| Equipment to Output AC Current (at the listed frequencies) ^{FO} | | | |
| 1 Hz to 10 Hz | 200 mA to 1.999 900 A | 620 μ A/A + 200 μ A | |
| 10Hz to 10 kHz | 200 mA to 1.999 900 A | 735 μ A/A + 200 μ A | |
| 10 kHz to 30 kHz | 200 mA to 1.999 900 A | 3 mA/A + 200 μ A | |
| Equipment to Output AC Current (at the listed frequencies) ^{FO} | | | |
| 1 Hz to 10 Hz | 2 A to 19.999 00 A | 820 μ A/A + 2 mA | |
| 10Hz to 10 kHz | 2 A to 19.999 00 A | 2.5 mA/A + 2 mA | |
| Temperature Calibration, Indication, and Control Equipment used with Thermocouple (Type B) ^{FO} | 600 °C to 800 °C | 0.51 °C | Electrical Simulation of Thermocouple Output Fluke 5520A |
| | 800.01 °C to 1 000 °C | 0.4 °C | |
| | 1 000.01 °C to 1 550 °C | 0.35 °C | |
| | 1 550.01 °C to 1 820 °C | 0.39 °C | |
| Temperature Calibration, Indication, and Control Equipment used with Thermocouple (Type C) ^{FO} | 0 °C to 150 °C | 0.35 °C | |
| | 150.01 °C to 650 °C | 0.3 °C | |
| | 650.01 °C to 1 000 °C | 0.36 °C | |
| | 1 000.01 °C to 1 800 °C | 0.58 °C | |
| | 1 800.01 °C to 2 316 °C | 0.97 °C | |



Certificate of Accreditation: Supplement

PCI

8100 Brownleigh Dr., #100A, Raleigh, NC 27617
 Contact Name: Frank Reagan Phone: 919-781-7787

Accreditation is granted to the facility to perform the following calibrations:

Electrical

| MEASURED INSTRUMENT, QUANTITY OR GAUGE | RANGE (AND SPECIFICATION WHERE APPROPRIATE) | CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm) | CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED |
|--|---|--|---|
| Temperature Calibration, Indication, and Control Equipment used with Thermocouple (Type E) ^{FO} | -250 °C to -100 °C | 0.58 °C | Electrical Simulation of Thermocouple Output Fluke 5520A |
| | -99.99 °C to -25 °C | 0.19 °C | |
| | -24.99 °C to 350 °C | 0.17 °C | |
| | 350.01 °C to 650 °C | 0.19 °C | |
| | 650.01 °C to 1 000 °C | 0.25 °C | |
| Temperature Calibration, Indication, and Control Equipment used with Thermocouple (Type J) ^{FO} | -210 °C to -100 °C | 0.32 °C | |
| | -99.99 °C to -30 °C | 0.21 °C | |
| | -29.99 °C to 150 °C | 0.17 °C | |
| | 150.01 °C to 760 °C | 0.20 °C | |
| | 760.01 °C to 1200 °C | 0.27 °C | |
| Temperature Calibration, Indication, and Control Equipment used with Thermocouple (Type K) ^{FO} | -200 °C to -100 °C | 0.39 °C | |
| | -99.99 °C to -25 °C | 0.21 °C | |
| | -24.99 °C to 120 °C | 0.19 °C | |
| | 120.01 °C to 1 000 °C | 0.3 °C | |
| | 1 000.01 °C to 1 372 °C | 0.47 °C | |
| Temperature Calibration, Indication, and Control Equipment used with Thermocouple (Type N) ^{FO} | -200 °C to -100 °C | 0.47 °C | Fluke 5520A |
| | -99.99 °C to -25 °C | 0.26 °C | |
| | -24.99 °C to 120 °C | 0.22 °C | |
| | 120 .01°C to 410 °C | 0.21 °C | |
| | 410.01 °C to 1 300 °C | 0.32 °C | |
| Temperature Calibration, Indication, and Control Equipment used with Thermocouple (Type R) ^{FO} | 0 °C to 250 °C | 0.66 °C | |
| | 250.01 °C to 400 °C | 0.41 °C | |
| | 400.01°C to 1 000 °C | 0.39 °C | |
| | 1 000 .01 °C to 1 767 °C | 0.47 °C | |
| Temperature Calibration, Indication, and Control Equipment used with Thermocouple (Type S) ^{FO} | 0 °C to 250 °C | 0.55 °C | |
| | 250.01 °C to 1 000 °C | 0.42 °C | |
| | 1 000.01 °C to 1 400 °C | 0.43 °C | |
| | 1 400.01 °C to 1 767 °C | 0.54 °C | |
| Temperature Calibration, Indication, and Control Equipment used with Thermocouple (Type T) ^{FO} | -250 °C to -150 °C | 0.73 °C | |
| | -149.99 °C to 0 °C | 0.28 °C | |
| | 0 °C to 120 °C | 0.19 °C | |
| | 120.01 °C to 400 °C | 0.17 °C | |



Certificate of Accreditation: Supplement

PCI

8100 Brownleigh Dr., #100A, Raleigh, NC 27617
 Contact Name: Frank Reagan Phone: 919-781-7787

Accreditation is granted to the facility to perform the following calibrations:

Electrical

| MEASURED INSTRUMENT, QUANTITY OR GAUGE | RANGE (AND SPECIFICATION WHERE APPROPRIATE) | CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm) | CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED |
|--|---|---|---|
| Temperature Calibration, Indication, and Control Equipment used with Thermocouple (Type U) ^{FO} | -200 °C to 0 °C | 0.65 °C | Fluke 5520A |
| | 0.01 °C to 600 °C | 0.32 °C | |
| Temperature Calibration, Indication, and Control Equipment used with RTD (Pt 385) 100 Ω ^{FO} | -200 °C to 0 °C | 0.05 °C | |
| | 0 °C to 100 °C | 0.07 °C | |
| | 100 °C to 300 °C | 0.09 °C | |
| | 300 °C to 400 °C | 0.1 °C | |
| | 400 °C to 630 °C | 0.12 °C | |
| Temperature Calibration, Indication, and Control Equipment used with RTD (Pt 3926) 100 Ω ^{FO} | -200 °C to 0 °C | 0.05 °C | |
| | 0 °C to 100 °C | 0.07 °C | |
| | 100 °C to 300 °C | 0.09 °C | |
| | 300 °C to 400 °C | 0.1 °C | |
| Temperature Calibration, Indication, and Control Equipment used with RTD (Pt 3916) 100 Ω ^{FO} | 400 °C to 630 °C | 0.12 °C | Electrical Simulation of RTD Output Fluke 5500A- SC300 Calibrator |
| | -200 °C to -190 °C | 0.25 °C | |
| | -190 °C to -80 °C | 0.04 °C | |
| | -80 °C to 0 °C | 0.05 °C | |
| | 0 °C to 100 °C | 0.06 °C | |
| | 100 °C to 260 °C | 0.07 °C | |
| | 260 °C to 300 °C | 0.09 °C | |
| | 300 °C to 400 °C | 0.1 °C | |
| | 400 °C to 600 °C | 0.12 °C | |
| 600 °C to 630 °C | 0.23 °C | | |
| Temperature Calibration, Indication, and Control Equipment used with RTD (Pt 385) 200 Ω ^{FO} | -200 °C to 100 °C | 0.04 °C | |
| | 100 °C to 260 °C | 0.05 °C | |
| | 260 °C to 300 °C | 0.12 °C | |
| | 300 °C to 400 °C | 0.13 °C | |
| | 400 °C to 600 °C | 0.14 °C | |
| | 600 °C to 630 °C | 0.16 °C | |



Certificate of Accreditation: Supplement

PCI

8100 Brownleigh Dr., #100A, Raleigh, NC 27617
 Contact Name: Frank Reagan Phone: 919-781-7787

Accreditation is granted to the facility to perform the following calibrations:

Electrical

| MEASURED INSTRUMENT, QUANTITY OR GAUGE | RANGE (AND SPECIFICATION WHERE APPROPRIATE) | CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm) | CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED |
|---|---|---|---|
| Temperature Calibration, Indication, and Control Equipment used with RTD (Pt 385) 500 Ω ^{FO} | -200 °C to -80 °C | 0.04 °C | Electrical Simulation of RTD Output Fluke 5500A- SC300 Calibrator |
| | -80 °C to 100 °C | 0.05 °C | |
| | 100 °C to 260 °C | 0.06 °C | |
| | 260 °C to 400 °C | 0.08 °C | |
| | 400 °C to 600 °C | 0.09 °C | |
| | 600 °C to 630 °C | 0.11 °C | |
| Temperature Calibration, Indication, and Control Equipment used with RTD (Pt 385) 1 000 Ω ^{FO} | -190 °C to 0 °C | 0.03 °C | |
| | 0 °C to 100 °C | 0.04 °C | |
| | 100 °C to 260 °C | 0.05 °C | |
| | 260 °C to 300 °C | 0.06 °C | |
| | 300 °C to 600 °C | 0.07 °C | |
| | 600 °C to 630 °C | 0.23 °C | |
| Temperature Calibration, Indication, and Control Equipment used with RTD (Ni 385) 120 Ω ^{FO} | -80 °C to 100 °C | 0.08 °C | |
| | 100 °C to 260 °C | 0.14 °C | |
| Temperature Calibration, Indication, and Control Equipment used with RTD (Cu 427) 10 Ω ^{FO} | -100 °C to 260 °C | 0.3 °C | Fluke 5500A-SC300 Calibrator |
| Volts ^{FO} | 1.8 mV to 105 V pk-pk | 2.5 mV/V + 100 μ V | |
| Leveled Sine Wave - Flatness ^{FO} | 50 kHz Reference | 20 mV/V + 200 μ V | |
| | 50 kHz to 100 MHz | 15 mV/V + 100 μ V | |
| | 100 MHz to 300 MHz | 20 mV/V + 100 μ V | |
| Time Markers ^{FO} | 2 ns to 1 μ s | 0.002 5 % of reading | |
| | 2 μ s to 50 μ s | 0.002 5 % of reading + 0.015 T | |
| | 100 μ s to 5 s | 0.002 5 % of reading + 0.001 T | |
| Equipment to Measure DC Current Toroidal Clamp Meter ^{FO} | 1.65 to 16.5 ADC | 0.004 6 A + 0.002 77 A / A | Fluke 5520A with Fluke 50-turn coil |
| | 16.5 to 55 ADC | 0.017 A + 0.002 9 A / A | |
| | 55 to 150 ADC | 0.018 A + 0.002 92 A / A | |
| | 150 to 550 ADC | 0.073 + 0.002 87 A / A | |
| | 550 to 1 000 ADC | 0.07 A + 0.003 11 A / A | |



Certificate of Accreditation: Supplement

PCI

8100 Brownleigh Dr., #100A, Raleigh, NC 27617
 Contact Name: Frank Reagan Phone: 919-781-7787

Accreditation is granted to the facility to perform the following calibrations:

Electrical

| MEASURED INSTRUMENT, QUANTITY OR GAUGE | RANGE (AND SPECIFICATION WHERE APPROPRIATE) | CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm) | CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED |
|--|---|---|---|
| Equipment to Measure AC Current Toroidal Clamp Meter ^{FO} | 1.65 AAC to 16.5 AAC 45 Hz to 65 Hz | 0.005 8 A + 0.003 27 A / A | Fluke 5520A with Fluke 50-turn coil |
| | 1.65 AAC to 16.5 AAC 65 Hz to 100 Hz | 0.004 7 A + 0.009 11 A / A | |
| | 1.65 AAC to 16.5 AAC 100 Hz to 440 Hz | 0.006 2 A + 0.009 33 A / A | |
| | 16.5 AAC to 150 AAC 45 Hz to 65 Hz | 0.031 A + 0.003 52 A / A | |
| | 16.5 AAC to 150 AAC 65 Hz to 100 Hz | 0.055 A + 0.009 23 A / A | |
| | 16.5 AAC to 150 AAC 45 Hz to 440 Hz | 0.055 A + 0.009 73 A / A | |
| | 150 AAC to 1 000 AAC 45 Hz to 65 Hz | 0.14 A + 0.003 51 A / A | |
| | 150 AAC to 1 000 AAC 65 Hz to 100 Hz | 0.14 A + 0.009 22 A / A | |
| | 150 AAC to 1000 AAC 100 Hz to 440 Hz | 0.13 A + 0.009 19 A / A | |
| Equipment to Measure DC Current Non-Toroidal Clamp Meter ^{FO} | 1.65 ADC to 16.5 ADC | 0.024 A + 0.005 75 A / A | |
| | 16.5 ADC to 55 ADC | 0.16 + 0.0057 8 A / A | |
| | 55 ADC to 150 ADC | 0.16 A + 0.005 79 A / A | |
| | 150 ADC to 550 ADC | 0.58 A + 0.005 8 A / A | |
| | 550 ADC to 1 000 ADC | 0.58 A + 0.005 89 A / A | |
| Equipment to Measure AC Current Non-Toroidal Clamp Meter ^{FO} | 1.65 AAC to 16.5 AAC 45 Hz to 65 Hz | 0.035 A + 0.006 51 A / A | |
| | 1.65 AAC to 16.5 AAC 65 Hz to 100 Hz | 0.035 A + 0.011 6 A / A | |
| | 1.65 AAC to 16.5 AAC 100 Hz to 440 Hz | 0.035 A + 0.011 8 A / A | |
| | 16.5 AAC to 150 AAC 45 Hz to 65 Hz | 0.29 A + 0.006 6 A / A | |
| | 16.5 AAC to 150 AAC 65 Hz to 100 Hz | 0.29 A + 0.01 16 A / A | |
| | 16.5 AAC to 150 AAC 100 Hz to 440 Hz | 0.29 A + 0.012 1 A / A | |
| | 150 AAC to 1 000 AAC 45 Hz to 65 Hz | 1.0 A + 0.006 61 A / A | |
| | 150 AAC to 1 000 AAC 65 Hz to 100 Hz | 1.0 A + 0.011 6 A / A | |
| | 150 AAC to 1 000 AAC 100 Hz to 440 Hz | 0.99 A + 0.016 3 A / A | |



Certificate of Accreditation: Supplement

PCI

8100 Brownleigh Dr., #100A, Raleigh, NC 27617
 Contact Name: Frank Reagan Phone: 919-781-7787

Accreditation is granted to the facility to perform the following calibrations:

Mechanical

| MEASURED INSTRUMENT, QUANTITY OR GAUGE | RANGE (AND SPECIFICATION WHERE APPROPRIATE) | CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm) | CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED |
|---|--|--|---|
| Equipment to Measure Pressure ^F | -14.9 psig to -0.05 psig | 0.0012 psig/psig + 0.000 00 376 psig | Mensor CPC6050 |
| | -30 inH ₂ O to -10 inH ₂ O | 0.000 11 inH ₂ O/ inH ₂ O + 0.000 007 3 inH ₂ O | Fluke 7250lp |
| | -10 inH ₂ O to 10 inH ₂ O | 0.000 095 inH ₂ O/ inH ₂ O + 0.000 13 inH ₂ O | |
| | 10 inH ₂ O to 30 inH ₂ O | 0.000 11 inH ₂ O/ inH ₂ O + 0.000 007 3 inH ₂ O | |
| | 1 psig to 10 psig | 0.001 2 psig + 0 000 001 08 psig/psig | Mensor CPC6050 |
| | 10 psig to 20 psig | 0.000 21 psig + 0 000 117 psig/psig | Fluke 2700G-BG200K |
| | 20 psig to 25 psig | 0.003 6 + 0.000 001 31 psig/psig | |
| | 25 psig to 500 psig | 0.000 16 psig/psig + 0.000 72 psig | |
| | 500 psig to 1 000 psig | 0.000 13 psig + 0.000 118 psig/psig | Mensor CPC6050 |
| Equipment to Measure Pressure ^{FO} | 1 000 psig to 2 000 psig | 0.000 007 4 psig + 0.000 119 psig/psig | Fluke PM600-A14M |
| | 2 000 psig to 5 000 psig | 0.57 + 0.000 001 01 psig/psig | Fluke 2700G-G35M |
| | 5 000 psig to 15 000 psig | 0.1 % of reading | Ametek T-150 |
| Equipment to Measure Pressure ^O | -15 psig to -0.01 psig | 0.009 3 - 0.000 000 254 psig/psig | Fluke 2700G-BG200K |
| | 0.01 psig to 30 psig | 0.003 6 + 0.000 001 31 psig/psig | |
| | 30 psig to 90 psig | 0.011 psig | Fluke PM600-A2M |
| | 90 psig to 300 psig | 0.002 7 + 0.000 106 psig/psig | Fluke PM600-A2M |
| | 300 psig to 1 000 psig | 0.000 001 9 + 0.000 115 psig/psig | Fluke PM600-A7M |
| | 1 000 psig to 2 000 psig | -0.000 68 + 0.000 116 psig/psig | Fluke PM600-A14M |
| Torque Wrenches ^{FO} | 0.5 lbf•in to 5 lbf•in | 1 % of reading | Mountz Indicator PTT-2000 w/ BMX80z Transducer |
| | 5 lbf•in to 50 lbf•in | 1 % of reading | Mountz Indicator PTT-2000 w/ BMX50I Transducer |
| | 50 lbf•in to 500 lbf•in | 1 % of reading | Mountz Indicator PTT-2000 w/ BMX500I Transducer |
| | 50 lbf•ft to 250 lbf•ft | 1 % of reading | Mountz Indicator PTT-2000 w/ BMX250F Transducer |



Certificate of Accreditation: Supplement

PCI

8100 Brownleigh Dr., #100A, Raleigh, NC 27617
 Contact Name: Frank Reagan Phone: 919-781-7787

Accreditation is granted to the facility to perform the following calibrations:

Mechanical

| MEASURED INSTRUMENT, QUANTITY OR GAUGE | RANGE (AND SPECIFICATION WHERE APPROPRIATE) | CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm) | CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED | |
|--|---|--|--|--|
| Pipettes and Repeaters ^{FO} | 0.1 μ L to 2 μ L | 0.020 8 μ L/ μ L + 0.017 μ L | Photometric using Artel PCS TM | |
| | 2 μ L to 10 μ L | 0.001 23 μ L/ μ L + 0.056 μ L | | |
| | 10 μ L to 30 μ L | 0.002 95 μ L/ μ L + 0.038 μ L | | |
| | | 20 μ L to 200 μ L | 0.002 04 μ L/ μ L + 0.089 μ L | Gravimetric Tare Addition with Moisture Trap |
| | | 200 μ L to 1 000 μ L | 0.001 58 μ L/ μ L + 0.18 μ L | |
| | | 1 000 μ L to 5 000 μ L | 0.001 79 μ L/ μ L + 0.76 μ L | |
| | | 5 000 μ L to 10 000 μ L | 0.006 95 μ L/ μ L + 25 μ L | |
| | | 10 000 μ L to 20 000 μ L | 0.006 02 μ L/ μ L + 16 μ L | |
| Equipment to Measure Flow ^F | 2 sccm to 20 sccm | 0.003 2 sccm/sccm + 0.000 29 sccm | Fluke Molbloc+ System | |
| | 20 sccm to 200 sccm | 0.003 1 sccm/sccm + 0.057 sccm | | |
| | 300 sccm to 3 000 sccm | 0.000 55 sccm/sccm + 4.2 sccm | | |
| | 3 slm to 50 slm | 0.001 9 lm/lm + 0.002 5 lm | | |
| | 50 slm to 100 slm | 0.003 2 slm/slm + 0.000 16 slm | | |

Time & Frequency

| MEASURED INSTRUMENT, QUANTITY OR GAUGE | RANGE (AND SPECIFICATION WHERE APPROPRIATE) | CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm) | CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED |
|--|---|--|---|
| Equipment to measure Time and Frequency ^F | 10 MHz | 2 parts in 10^{-12} | Quartzlock E8-X GPS Reference |
| | 1 PPS | 2 parts in 10^{-12} | |
| | Up to 20 MHz | 1 part in 10^{-10} | HP 3325A with GPS |
| | 20 MHz to 300 MHz | 2 parts in 10^{-10} | Fluke 5500A monitored with GPS disciplined Fluke PM6681 |
| Equipment to measure Time and Frequency ^O | 0.001 Hz to 300 MHz | 1.2×10^{-5} Hz/Hz | Fluke PM6681 |
| Equipment to output Time and Frequency ^{FO} | 0.001 Hz to 300 MHz | 1.2×10^{-5} Hz/Hz | Fluke PM6681 |
| Equipment to source Time and Frequency ^F | Up to 300 MHz | 2 parts in 10^{-10} | GPS Disciplined Fluke PM6681 |
| Stopwatches /Timers ^{FO} | Up to 48 hours | 0.05 sec/day | Timometer 4500 |
| | 60 sec to 8 640 0 sec | 0.038 sec/day | Fluke PM6681 NIST 960-12 Totalized Method |



Certificate of Accreditation: Supplement

PCI

8100 Brownleigh Dr., #100A, Raleigh, NC 27617
 Contact Name: Frank Reagan Phone: 919-781-7787

Accreditation is granted to the facility to perform the following calibrations:

Chemical

| MEASURED INSTRUMENT, QUANTITY OR GAUGE | RANGE (AND SPECIFICATION WHERE APPROPRIATE) | CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm) | CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED |
|--|---|--|--|
| Conductivity Meters ^F | 18.18 M Ω -cm | 0.27 M Ω -cm | Thornton 200CRw/ 240-201 Conductivity |

Thermodynamic

| MEASURED INSTRUMENT, QUANTITY OR GAUGE | RANGE (AND SPECIFICATION WHERE APPROPRIATE) | CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm) | CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED |
|---|---|--|--|
| Temperature Measurement PRT ^F | -196 °C | 0.01 °C | Hart 1575A, Fluke 5699, EDL Cryo-Cal |
| | -80 °C to 0 °C | 0.007 4 °C + 0.000 088 °C/°C | Hart 1575A, Fluke 5699 and Fluke 7381 |
| | 0 °C to 110 °C | 0.007 4 °C + 0.000 028 3 °C/°C | Hart 1575A, Fluke 5699 and Fluke 7341 |
| | 110 °C to 150 °C | 0.011 °C + 0.000 040 4 °C/°C | Hart 1575A, Fluke 5699 and Fluke 7341 |
| | 150 °C to 200 °C | 0.026 °C + 0.000 111 °C/°C | Hart 1575A, Fluke 5699 and Fluke 6102 |
| | 200 °C to 550 °C | 0.002 8 °C + 0.000 035 °C/°C | Hart 1575, Fluke 5699 and Ultra Bath |
| Temperature Measurement PRT ^O | -196 °C to 660 °C | 0.025 °C | Hart 1575 and Fluke 5699 |
| Temperature Measurement Thermistor ^{FO} | -40 °C to 140 °C | 0.025 °C | Hart 1575 and Hart 5699 SPRT |
| Temperature Measurement LIG Thermometer ^F | -80 °C to 500 °C | 0.025 °C | Hart 2560 and Hart 5628 SPRT |
| Temperature Measurement Thermocouple – Type T ^F | -196°C to 400 °C | 0.25 °C | |
| Temperature Measurement Thermocouple Types J & K ^F | 0°C to 600 °C | 0.35 °C | |
| Equipment to Measure Temperature ^F | 0.01 °C | 0.005 °C | Triple Point of Water |
| | -38.834 4 °C | 0.001 °C | ISO Tech Fixed Cell - Mercury |
| | 156.598 5 °C | 0.002 °C | ISO Tech Fixed Cell – Indium |
| | 419.527 °C | 0.002 °C | ISO Tech Fixed Cell - Zinc |
| Equipment to Measure Humidity ^F | 10 % RH to 95 % RH | 0.5 % RH | Thunder Scientific 2500 Humidity |



Certificate of Accreditation: Supplement

PCI

8100 Brownleigh Dr., #100A, Raleigh, NC 27617
Contact Name: Frank Reagan Phone: 919-781-7787

Accreditation is granted to the facility to perform the following calibrations:

1. The CMC (Calibration and Measurement Capability) stated for calibrations included on this scope of accreditation represents the smallest measurement uncertainty attainable by the laboratory when performing a more or less routine calibration of a nearly ideal device under nearly ideal conditions. It is typically expressed at a confidence level of 95 % using a coverage factor k (usually equal to 2). The actual measurement uncertainty associated with a specific calibration performed by the laboratory will typically be larger than the CMC for the same calibration since capability and performance of the device being calibrated and the conditions related to the calibration may reasonably be expected to deviate from ideal to some degree.
2. The laboratories range of calibration capability for all disciplines for which they are accredited is the interval from the smallest calibrated standard to the largest calibrated standard used in performing the calibration. The low end of this range must be an attainable value for which the laboratory has or has access to the standard referenced. Verification of an indicated value of zero in the absence of a standard is common practice in the procedure for many calibrations but by its definition it does not constitute calibration of zero capacity.
3. The presence of a superscript F means that the laboratory performs calibration of the indicated parameter at its fixed location. Example: Outside Micrometer^F would mean that the laboratory performs this calibration at its fixed location.
4. The presence of a superscript FO means that the laboratory performs calibration of the indicated parameter both at its fixed location and onsite at customer locations. Example: Outside Micrometer^{FO} would mean that the laboratory performs this calibration at its fixed location and onsite at customer locations.
5. Measurement uncertainties obtained for calibrations performed at customer sites can be expected to be larger than the measurement uncertainties obtained at the laboratories fixed location for similar calibrations. This is due to the effects of transportation of the standards and equipment and upon environmental conditions at the customer site which are typically not controlled as closely as at the laboratories fixed location.
6. The term T represents time in s (seconds).